

Certificate of Compliance

Certificate: 1383315

Master Contract: 204767

Project: 1383315


Date Issued: January 6, 2003

Issued to: Camille Bauer AG
Aaraustrasse 7
CH-5610 Wohlen
SWITZERLAND
Attention: Mr. Daniel Huber


*The products listed below are eligible to bear the CSA Mark shown
with adjacent indicators 'C' and 'US'*



Issued by:


Eknath Desai

Authorized by:


Ray Fadavi, P.Eng; MBA
Operations Manager

PRODUCTS

CLASS 3631 04 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 84 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - CERTIFIED TO U.S. STANDARDS

PART A - AC Voltage Transducer, Model SINEAX U539, permanently connected, rated 85-230V AC/DC; 40-400 Hz or 24-60V AC/DC, 3VA, Equipment Class II, Pollution Degree 2, Installation Category III.

PART B - AC Current Transducer, Model SINEAX I538, permanently connected, rated 85-230V AC/DC OR 24-60V AC/DC, 3VA, Equipment Class II, Pollution Degree 2, Installation Category III.

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



CSA INTERNATIONAL

Certificate: 1383315

Master Contract: 204767

Project: 1383315

Date: January 6, 2003

APPLICABLE REQUIREMENTS

- CAN/CSA-C22.2 No. 1010.1-92 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (Includes Amendment 1)
- CAN/CSA-C22.2 No. 1010.1B-97 - Amendment 2 to CAN/CSA-C22.2 No. 1010.1-92, "Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements"
- UL Std No. 3111.1 - Equipment for Measurement Use; Part 1: General Requirements



Supplement to Certificate of Compliance

Certificate: 1383315

Master Contract: 204767

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
1383315	January 06, 2003	Original Certification.